

NanoMAXX DICOM Conformance Statement

This document contains confidential information that is proprietary to SonoSite. Neither the document nor the information contained therein should be disclosed or reproduced in whole or in part, without express written consent of SonoSite.

Document Number: D07914

Revision: C

Copyright © 2016 FUJIFILM SonoSite, Inc. All rights reserved

CONTENTS

CHAPTER 1	1
INTRODUCTION	1
About this document.....	1
Changes in this version	1
Conformance Statement Overview	1
DICOM Background.....	2
Definitions	3
Reference Documents.....	4
CHAPTER 2	5
IMPLEMENTATION MODEL	5
Application Data Flow Diagram	6
Exam Archive.....	7
Get Status.....	8
Patient.....	8
Exam Export.....	8
Functional Definition of AE's.....	8
Store	8
Worklist	9
Sequencing of Real-World Activities	9
CHAPTER 3	11
AE SPECIFICATIONS	11
Store AE Specification.....	11
Association Establishment Policies.....	11
Association Initiation by Real-World Activity	12
Proposed Presentation Contexts to an Archiver.....	12
Common Composite Image IOD Modules.....	15
Store AE Behavior to C Store Status.....	23
Modality Worklist AE Specification.....	24
Association Establishment Practices	24
Association Initiation by Real-World Activity	25
Proposed Presentation Contexts to a Worklist Server.....	26
Modality Worklist Attributes.....	26
Worklist AE Behavior to C-FIND Status	29
Worklist AE Behavior to C-FIND Status	29
Media Export AE-Specification.....	30
Implementation Model.....	30
AE Specifications	31

Augmented and Private Application Profiles	32
Media Configuration.....	32
Media Storage SOP Class.....	32
Information Module Definitions.....	33
CHAPTER 4	37
COMMUNICATION PROFILES	37
TCP/IP Stack.....	37_Toc442366506
CHAPTER 5	39
EXTENSIONS/SPECIALIZATIONS/PRIVATIZATIONS	39
Standard Extended/Specialized/Private SOPs	39
Private Transfer Syntaxes.....	39
CHAPTER 6	41
CONFIGURATION	41
AE Title/Presentation Address Mapping.....	41
Configurable Parameters.....	41
NanoMAXX Configurable Parameters per Network Location.....	41
Configurable Parameters per Remote Device Instance	41
Other Configurable Parameters.....	42
CHAPTER 7	43
SUPPORT OF EXTENDED CHARACTER SETS.....	43

Introduction

About this document

This document describes the SonoSite NanoMAXX[®] Ultrasound System’s conformance to the ACR-NEMA DICOM (Digital Imaging and Communications in Medicine) standard and satisfies the DICOM requirement for a vendor conformance specification.

The NanoMAXX system is an ultrasound imaging device. The DICOM options of the NanoMAXX system provide a means to query the Information System for scheduled procedures using Modality Worklist and send images to storage servers and removable USB media.

Throughout this document DICOM storage servers will be referred to as archivers. For a device to be classified as an archiver it must be capable of receiving DICOM store commands. Archivers are primarily comprised of PACS.

This document is written with respect to ACR-NEMA DICOM version number 3.0 -2007.

Changes in this version

Revision	Description of Change
C	Add C-Store and Modality Worklist
B	Updated to clarify specifics on private tag inclusions
A	Initial release

Conformance Statement Overview

The NanoMAXX Ultrasound System implements the necessary DICOM services to download work lists from an information system, save acquired images to a network storage, and store DICOM files onto removable media.

Table 1-1 provides an overview of the network services supported by the NanoMAXX Ultrasound System.

Table 1-1: Networking Services

Networking SOP Classes	User of Service (SCU)	Provider of Service (SCP)
Transfer		
Ultrasound Image Storage	Yes	No
Ultrasound Image Storage (Retired)	Yes	No
Secondary Capture Image Storage	Yes	No

Table 1-1: Networking Services

Networking SOP Classes	User of Service (SCU)	Provider of Service (SCP)
Workflow Management		
Modality Worklist	Yes	No
General		
Verification	Yes	No

Table 1-2 provides an overview of the media storage services supported by the NanoMAXX Ultrasound System.

Table 1-2: Media Storage Services

SOP Classes	Role
Media Storage Directory Storage	FSC
Ultrasound Image Storage	FSC
Secondary Capture Image Storage	FSC

DICOM Background

The DICOM information exchange specification provides a definitive structure of commands and information that allow for the inter-communication of medical imaging devices. Developed by the American College of Radiology (ACR) and the National Electrical Manufacturers Association (NEMA), the DICOM standard strives to promote communication of image information through the use of a standardized set of command classes and information semantics.

The DICOM standard defines classes of information that are common to many modalities of medical imaging. However, to meet the specific needs of information content for such a diverse range of information, the DICOM specification defines structures for a multitude of medical data. To alleviate the need for applications to implement every aspect of the DICOM specification, a list of conformance tables for every modality was created to define the minimum set of information necessary for data exchanges. A requirement of the DICOM specification is to maintain a compliance document that outlines a subset of DICOM services and data classes that are supported by a device. The purpose of this document is to define a subset of DICOM for the exchange of information with the SonoSite NanoMAXX via its DICOM feature.

Definitions

AE	Application Entity
ANSI	American National Standards Institute
CW	Continuous Wave
DICOM	Digital Imaging and Communications in Medicine
DIMSE	DICOM Message Service Element
FSC	File Set Creator
HIS	Hospital Information System
IE	Information Entity
IOD	Information Object Definition
KHz	Kilohertz
LUT	Look Up Table
MPPS	Modality Performed Procedure Step
PACS	Picture Archive and Communication System
PW	Pulsed Wave
PDU	Protocol Data Unit
PPS	Performed Procedure Step
RGB	Red, Green, Blue
RIS	Radiology Information System
SC	Secondary Capture
SCU	Service Class User (Client)
SCP	Service Class Provider (Server)
SOP	Service Object Pair
SPS	Scheduled Procedure Step
TCP/IP	Transmission Control Protocol/Internet Protocol
UID	Unique Identifier
US	Ultrasound
USB	Universal Serial Bus
UTC	Coordinated Universal Time
VOI	Value Of Interest
VR	Value Representation

Reference Documents

ACR-NEMA DICOM Standard Version 3.0 – 2007.

CHAPTER 2

Implementation Model

The NanoMAXX DICOM feature incorporates the DICOM 3.0 standard for networked image storage and Modality Worklist functions. Scheduled Procedures are queried from the HIS/RIS Worklist SCP and presented to the operator for selection. Performed Procedures¹ are transferred from the NanoMAXX ultrasound system using standard network connections to be stored on a DICOM compatible archiver.

The behavior of how images are sent depends on which Transfer Images setting is selected during DICOM setup of locations. Two selections are offered, “During the exam” (in-progress transfer mode) or “End of exam” (batch transfer mode).

For batch transfer mode, NanoMAXX allows up to four archivers and one Worklist server to be selected at any given time. The devices are selected using DICOM Setup mode with all selected devices being placed into a destination list.

For in-progress transfer mode, NanoMAXX allows only one archive device and one Worklist server to be selected at any given time. Like batch transfer mode, the devices are selected using DICOM Setup mode with all selected devices annotated in the destination list.

A maximum of 200 Scheduled Procedures Steps may be queried from the selected Modality Worklist SCP. The Worklist is persisted to non-volatile memory so that it can be accessed during portable exams. Both manual and automatic queries are supported. Automatic queries are user configurable and are performed in the background at periodic intervals.

During an exam all saved images are written to internal storage. For batch transfer mode, when the exam completes, all images associated with it are marked as Archive Pending for transfer to each device in the current destination list. If a network connection is present then transfer begins immediately.

Performed Procedures are Archived to devices in the destination list sequentially, starting with the first selected archiver. Exam images are sent to each destination device in batch transfer mode; an association is opened, all exam images are transferred in acquisition order, and the association is closed. Once an exam is successfully transferred to a device then all images in the exam are marked as Archive Complete to that destination. Archiving then continues with the next device in the destination list. Once all devices in the destination list have successfully received each exam image then the Exam Archive is complete.

For in-progress transfer mode, images are transferred immediately after acquisition provided there is a network connection present.

Acquired images are sent to the destination archive device; an association is opened if closed, the acquired image is transferred and the association is left open at the end of image transfer. The association is left open in anticipation of transferring another acquired image. If another image is not acquired within nominally 30 seconds, then the association is closed to preserve network resources. Any additional images acquired during the exam are sent on a subsequent association(s) using the sequence described above.

¹ Performed Procedures consist of images from the ultrasound system.

When a Get Status is performed the current destination list is used in the same manner as with Exam Archive. For batch transfer mode, the devices are accessed sequentially, starting with the first selected archiver. For each device an association is opened, status is returned, and the association is closed. Status is obtained from archiver and Worklist devices using DICOM Verify (C-Echo). Once status is successfully returned from all devices in the destination list then Get Status is complete.

When a Get Status is performed during in-progress transfer mode, the archive device association is opened if closed, a C-Echo Request is issued and the C-Echo response status is reported. The association remains open while in DICOM Setup mode. Once DICOM Setup mode is exited the rules used for image acquisition apply to closing the association. This behavior allows an in-progress transfer with an open association to remain open through the Get Status process and allows subsequent image acquisition to be sent on the same association when acquisition constraints are met.


One or more completed exams may be selected from the exam list to have their images saved as DICOM files to the selected USB medium.

The user may choose to “Append” a completed exam. On the NanoMAXX system, this is treated as a new Series in the same Study as the original exam. The appended exam shows up as a separate line item in the Patient List form.

Application Data Flow Diagram

The diagram in [Figure 2-1](#) represents the relationship between the ultrasound system’s real-world activities (circles on the left), the local AE’s built into NanoMAXX (boxes in the center), and the remote AE’s built into the devices NanoMAXX communicates with using DICOM (boxes on the right).

Figure 2-1 Implementation Model



The following are the conditions that invoke real-world activities associated with AE's.

Exam Archive

- ▶ For batch transfer mode, End Exam with one or more images saved on internal storage. Exam end occurs when the Delta key is pressed while configured to End Exam, or when Patient setup mode is entered and End Exam is pressed or any of the patient fields are changed and the saves committed or when New is pressed and you exit Patient setup.
- ▶ For in-progress transfer mode after an image acquisition is complete. Image acquisition

occurs after the Save key is pressed.

- ▶ System startup with one or more images flagged as Archive Pending.

Get Status

- ❖ Operator Verify command in DICOM Setup mode.

Patient

- ❖ User enters Patient Setup screen, enters search criteria, and presses the Query key. The Worklist screen is entered and a list of matching Scheduled Procedures Steps are returned and displayed.

Exam Export

- ❖ User selects one or more completed exams from the exam list and the images for those exams are written to the selected removable media.

Functional Definition of AE's

Store

This AE handles sending ultrasound images to an archiver using the DICOM store SCU services.

Steps taken to Get Archiver Status:

```
A-ASSOCIATE
C-ECHO command
A-RELEASE
```

Steps taken to Send Exam to Archiver, batch transfer mode:

```
A-ASSOCIATE
for each exam image
{
    C-STORE Image SOP Instance
}
A-RELEASE
```

Steps taken to Send Exam to Archiver, in progress transfer mode:

```
A-ASSOCIATE
for each exam image acquired within timeout period AND not end of exam
{
    C-STORE Image SOP Instance
}
```

A-RELEASE

Worklist

This AE handles querying a Worklist SCP for a list of scheduled procedures using the DICOM Modality Worklist SCU services.

Steps taken to Get Worklist Status:

A-ASSOCIATE
C-ECHO command
A-RELEASE


Steps taken to Query a WorklistSCP:

A-ASSOCIATE
SEND C-FIND Requestcommand
{
 RECEIVE C-FIND Response
} While C-FIND status == pending AND response <= 200
A-RELEASE

Sequencing of Real-World Activities

All real world activities that initiate communication to remote AE's operate asynchronously with respect to each other and Workflow activities.


Figure 2-2 Sequencing Constraints - "During the Exam" Configuration



Under normal scheduled workflow conditions the sequencing constraints illustrated apply:

1. Worklist Query in initiated.
2. List of Scheduled Procedure Steps (SPS) are returned.
3. SPS item is selected from the Worklist and the Exam begins.
4. Image is acquired.
5. Association is opened with the Image Manager and the acquired image is stored. Subsequent image acquisitions are stored under the same association, if the acquisition completes within 30 seconds of the last Store operation. After 30 seconds of inactivity, the association is closed.

Figure 2-3 Sequencing Constraints - "End of Exam" Configuration



Under normal scheduled workflow conditions the sequencing constraints illustrated apply:

1. Worklist Query in initiated.
2. List of Scheduled Procedure Steps (SPS) are returned.
3. SPS item is selected from the Worklist and the Exam begins.
4. Images are acquired.
5. Exam is ended.
6. Image SOP instances acquired during the exam are stored to the Image Manager.

AE Specifications

Store AE Specification

The Store AE provides conformance to the following DICOM V3.0 SOP Classes as an SCU:

Table 3-1: Store AE SOP Class Support

SOP Class Name	SOP Class UID	Conformance Level
Verification	1.2.840.10008.1.1	Standard
Ultrasound Image Storage	1.2.840.10008.5.1.4.1.1.6.1	Standard
Ultrasound Image Storage (Retired)	1.2.840.10008.5.1.4.1.1.6	Standard
Secondary Capture Image Storage	1.2.840.10008.5.1.4.1.1.7	Standard

Association Establishment Policies

The Store AE will initiate an association to a device in response to the following real-world activities; Archive Exam or Image Acquisition, Review Archive, and GetStatus.

General

Maximum PDU size offered to SCP: 32,768 bytes

This is the maximum PDU size the Store AE can receive and is the value offered for the maximum PDU size in the Association Request command. Once the Association is open, if the Store AE receives a PDU that is larger than this value then the Association will be aborted.

Minimum PDU size accepted from SCP: 1,024 bytes

This is the minimum PDU size the Store AE can be configured to send. If the Store AE receives a maximum PDU size in the Association Accept response that is smaller than this value then the Association will be aborted immediately.

Maximum PDU size sent by SCU: 32,768 bytes

This is the maximum PDU size the Store AE can be configured to send. The maximum PDU size sent on any Store AE Association will be the smaller of the configured value and the maximum PDU size received in the Association Accept response.

Number of Associations

Number of simultaneous associations for the Store AE: 1

Asynchronous Nature

The Store AE will not use asynchronous operations.

Implementation Identifying Information

Implementation Class UID: "1.2.840.114340.3"

Implementation Version name: "Tiller_SV500"

Note

"114340" is registered by SonoSite with ANSI. Version name will be used initially as shown, but is subject to change with new versions of the DICOM capable application software.

Association Initiation by Real-World Activity

The Store AE will open associations to the storage devices listed in the current destination list in response to the following real-world activities; Archive Exam or Image Acquisition, Review Archive, and Get Status.

Association Initiation by: Archive Exam

The Archive Exam real-world activity if configured for batch transfer mode will cause the Store AE to open associations to each storage device listed in the current destination list.

Association Initiation by: Image Acquisition

The Image Acquisition real-world activity if configured for in-progress transfer mode will cause the Store AE to open an association to the selected storage device.

Association Initiation by: Review Archive

The Archive command real-world activity while in Review mode will cause the Store AE to open associations to each storage device listed in the current destination list.

Association Initiation by: Get Status

The Get Status real-world activity will cause the Store AE to open associations to each archiver listed in the current destination list.

Proposed Presentation Contexts to an Archiver

Table 3-2: Store AE Proposed Presentation Contexts to an Archiver

Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Extended Negotiation
Name	UID	Name List	UID List		
Verification	1.2.840.10008.1.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None

Table 3-2: Store AE Proposed Presentation Contexts to an Archiver

Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Extended Negotiation
Name	UID	Name List	UID List		
Ultrasound Image Storage	1.2.840.10008.5.1.4.1.1.6.1	Implicit VR Little Endian Explicit VR Little Endian JPEG Baseline (Process 1)*	1.2.840.10008.1.2 1.2.840.10008.1.2.1	SCU	None
Ultrasound Image Storage (Retired)	1.2.840.10008.5.1.4.1.1.6	Implicit VR Little Endian Explicit VR Little Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1	SCU	None
Secondary Capture Image Storage	1.2.840.10008.5.1.4.1.1.7	Implicit VR Little Endian Explicit VR Little Endian	1.2.840.10008.1.2 1.2.840.10008.1.2.1	SCU	None

* This Transfer Syntax is the only one proposed if JPEG Compression is configured.

Verification SOP Class

The Store AE provides standard conformance to the Verification SOP Class as an SCU. The remote SCP must support Verification in the same association as the Store Command (C-Store).

Ultrasound Image Storage SOP Class

The Ultrasound Image Storage SOP Class uses the Common Composite Image IOD Modules as shown in [Table 3-3](#).

Table 3-3: US Image IOD Modules

IE	Module	Reference	Usage
Patient	Patient	3.1.4.1	M
	Clinical Trial Subject	Not Used	U
Study	General Study	3.1.4.2	M
	Patient Study	3.1.4.3	U
	Clinical Trial Study	Not Used	U
Series	General Series	3.1.4.4	M
	Clinical Trial Series	Not Used	U

Table 3-3: US Image IOD Modules

IE	Module	Reference	Usage
Frame Of Reference	Frame Of Reference	Not Used	U
	Synchronization	Not Used	U
Equipment	General Equipment	3.1.4.5	M
Image	General Image	3.1.4.7	M
	Image Pixel	3.1.4.8	M
	Contrast/Bolus	Not Used	C
	Palette Color Lookup Table	Not Used	C
	US Region Calibration	3.1.4.9	U
	US Image	3.1.4.10	M
	Overlay Plane	Not Used	U
	VOI LUT	3.1.4.11	U
	SOP Common	3.1.4.12	M

Ultrasound Image Storage SOP Class (Retired)

The Ultrasound Image Storage SOP Class (Retired) uses the Common Composite Image IOD.

Secondary Capture Image Storage SOP Class

The Secondary Capture Image Storage SOP Class uses the Common Composite Image IOD Modules as shown in [Table 3-4](#).

Table 3-4: SC Image IOD Modules

IE	Module	Reference	Usage
Patient	Patient	3.1.4.1	M
	Clinical Trial Subject	Not Used	U
Study	General Study	3.1.4.2	M
	Patient Study	3.1.4.3	U
	Clinical Trial Study	Not Used	U
Series	General Series	3.1.4.4	M
	Clinical Trial Series	Not Used	U
Equipment	General Equipment	3.1.4.5	U
	SC Equipment	3.1.4.6	M
Image	General Image	3.1.4.7	M
	Image Pixel	3.1.4.8	M

Table 3-4: SC Image IOD Modules

IE	Module	Reference	Usage
	SC Image	Not Used	M
	Overlay Plane	Not Used	U
	Modality LUT	Not Used	U
	VOI LUT	3.1.4.11	U
	SOP Common	3.1.4.12	M

Common Composite Image IOD Modules

The section defines the Modules that are common to the Ultrasound, Ultrasound (Retired), and Secondary Capture Store SOP Classes.

Patient Module

Table 3-5 specifies the attributes used from the Patient Module. These attributes are used by the Ultrasound, Ultrasound (Retired), and Secondary Capture Image Storage SOP Classes. Attributes not listed are not used.

Table 3-21: Patient Module Attributes

Attribute Name	Tag	Type	Attribute Description
Patient's Name*	(0010,0010)	2	From Worklist or manually entered on Patient Setup screen (Last, First and Middle fields). All 5 Person Name Components are preserved when name comes from Worklist.
Patient ID*	(0010,0020)	2	From Worklist or manually entered on Patient Setup screen (ID field)
Patient's Birth Date*	(0010,0030)	2	From Worklist or manually entered on Patient Setup screen (Date of birth fields)
Patient's Sex*	(0010,0040)	2	From Worklist or manually entered on Patient Setup screen (Gender pick list)
Other Patient IDs	(0010,1000)	3	From Worklist
Ethnic Group	(0010,2160)	3	Manually entered on Patient Setup screen. Only sent for IMT Exam Types.

* This attribute cannot be modified by the user when coming from DICOM Worklist.

General Study Module

Table 3-6 specifies the attributes used from the General Study Module. These attributes are used by the Ultrasound, Ultrasound (Retired), and Secondary Capture Image Storage SOP Classes. Attributes not listed are not used.

Table 3-6: General Study Module Attributes

Attribute Name	Tag	Type	Attribute Description
Study Instance UID	(0020,000D)	1	From Worklist or automatically generated
Study Date	(0008,0020)	2	Procedure start date
Study Time	(0008,0030)	2	Procedure start time
Referring Physician's Name	(0008,0090)	2	From Worklist or manually entered on Patient Setup screen (Referring Dr. field) Note: Only last name component will be sent when manually entered.
Study ID	(0020,0010)	2	From Worklist (mapped from Requested Procedure ID attribute) or manually entered for unscheduled procedures. If no value is provided via worklist or manual entry, then a value will be automatically generated.
Accession Number ²	(0008,0050)	2	From Worklist or manually entered on Patient Setup screen (Accession field)
Study Description	(0008,1030)	3	From Worklist ¹ or selected manually on Patient Setup screen (Procedure Type pick list)
Referenced Study Sequence	(0008,1110)	3	From Worklist. Not sent if procedure was unscheduled.
>Referenced SOP Class UID	(0008,1150)	1C	
>Referenced SOP Instance UID	(0008,1155)	1C	
Procedure Code Sequence	(0008,1032)	3	Mapped from Worklist Requested Procedure Code Sequence, if performed. Not sent if procedure was unscheduled.
>Code Value	(0008,0100)	1C	
>Coding Scheme Designator	(0008,0102)	1C	
>Code Meaning	(0008,0104)	1C	
Name of Physician(s) Reading Study	(0008,1060)	3	Entered on Patient Setup screen. (Reading Dr. field) Note: Only last name component will be sent when manually entered.

¹Mapped from Scheduled Procedure Step Description (0040,0007) if it exists. Otherwise, Study Description is set to value of Requested Procedure Description (0032,1060) if it exists. If Requested Procedure Description (0032,1060) is also empty, Study Description is set to Requested Procedure Code Sequence (0032,1064) Code Meaning (0008,0104).

² This attribute cannot be modified by the user when coming from DICOM Worklist.

Patient Study Module

Table 3-7 specifies the attributes used from the Patient Study Module. These attributes are used by the Ultrasound, Ultrasound (Retired), and Secondary Capture Image Storage SOP Classes. Attributes not listed are not used.

Table 3-7: Patient Study Module Attributes

Attribute Name	Tag	Type	Attribute Description
Patient's Size	(0010,1020)	3	Only sent for Cardiac Exam types.
Patient's Weight	(0010,1030)	3	Only sent for Cardiac Exam types.
Additional Patient's History	(0010,21B0)	3	From Worklist or manually entered on Patient Setup screen (Indications field).

General Series Module

Table 3-8 specifies the attributes used from the General Series Module. These attributes are used by the Ultrasound, Ultrasound (Retired), and Secondary Capture Image Storage SOP Classes. Attributes not listed are not used.

Table 3-8: General Series Module Attributes

Attribute Name	Tag	Type	Attribute Description
Modality	(0008,0060)	1	"US"
Series Instance UID	(0020,000E)	1	Automatically generated
Series Number	(0020,0011)	2	" 1"
Laterality	(0020,0060)	2C	Zero Length
Series Date	(0008,0021)	3	Procedure start date
Series Time	(0008,0031)	3	Procedure start time
Protocol Name	(0018,1030)	3	Zero length
Series Description	(0008,103E)	3	Same as Performed Procedure Step Description
Operator's Name	(0008,1070)	3	Entered on Patient Setup screen (User field). The User's initials are transmitted in the last name component
Request Attributes Sequence	(0040,0275)	3	Only sent if the procedure originated from a Worklist Scheduled Procedure Step
>Requested Procedure ID	(0040,1001)	1C	From Worklist
>Scheduled Procedure Step ID	(0040,0009)	1C	From Worklist
>Scheduled Procedure Step Description	(0040,0007)	3	From Worklist

Table 3-8: General Series Module Attributes

Attribute Name	Tag	Type	Attribute Description
>Scheduled Protocol Code Sequence	(0040,0008)	3	From Worklist
>>Code Value	(0008,0100)	1C	
>>Coding Scheme Designator	(0008,0102)	1C	
>>Code Meaning	(0008,0104)	1C	
Performed Procedure Step ID	(0040,0253)	3	From Worklist (mapped from Scheduled Procedure Step ID) or Generated by NanoMAXX
Performed Procedure Step Start Date	(0040,0244)	3	Procedure start date
Performed Procedure Step Start Time	(0040,0245)	3	Procedure start time
Performed Procedure Step Description	(0040,0254)	3	Mapped from Worklist SPS description, if performed, or selected manually on Patient Setup screen (Procedure Type pick list)
Performed Protocol Code Sequence	(0040,0260)	3	Mapped From Worklist Scheduled Protocol Code sequence, if performed. Otherwise sent as zero length sequence.
>Code Value	(0008,0100)	1C	
>Coding Scheme Designator	(0008,0102)	1C	
>Code Meaning	(0008,0104)	1C	

General Equipment Module

Table 3-9 specifies the attributes used from the General Equipment Module. These attributes are used by the Ultrasound, Ultrasound (Retired), and Secondary Capture Image Storage SOP Classes. Attributes not listed are not used.

Table 3-9: General Equipment Module Attributes

Attribute Name	Tag	Type	Attribute Description
Manufacturer	(0008,0070)	2	"SonoSite, Inc."
Institution Name	(0008,0080)	3	Entered on Patient Setup screen(Institution)
Station Name	(0008,1010)	3	Host Name for current location
Manufacturer's Model Name	(0008,1090)	3	Model name (maps to product line)
Software Versions	(0018,1020)	3	ARM Firmware Version

SC Equipment Module

Table 3-10 describes the attributes used from the SC Equipment Module. These attributes are used by the Secondary Capture Image Storage SOP Class. Attributes not listed are not used.

Table 3-10: SC Equipment Module Attributes

Attribute Name	Tag	Type	Attribute Description
Conversion Type	(0008,0064)	1	“WSD”
Modality	(0008,0060)	3	“US”

General Image Module

Table 3-11 specifies the attributes used from the General Image Module. These attributes are used by the Ultrasound, Ultrasound (Retired), and Secondary Capture Image Storage SOP Classes. Attributes not listed are not used.

Table 3-11: General Image Module Attributes

Attribute Name	Tag	Type	Attribute Description
Instance Number	(0020,0013)	2	A number that identifies this image
Patient Orientation	(0020,0020)	2C	Zero Length
Content Date	(0008,0023)	2C	Image acquisition date
Content Time	(0008,0033)	2C	Image acquisition time
Derivation Description	(0008,2111)	3	“RGB to MONOCHROME2 conversion” - Sent for MONOCHROME2 images “RGB to JPEG Baseline 1 conversion” - Sent for JPEG Lossy compressed images.
Lossy Image Compression	(0028,2110)	3	01=Lossy Compressed - Only sent for MONOCHROME2 and JPEG Lossy Compressed images.
Lossy Image Compression Ratio	(0028,2112)	3	Set to 3 for MONOCHROME2 images. The approximate compression ratio is sent for JPEG Lossy Compressed images.

Image Pixel Module

Table 3-12 specifies the attributes used from the Image Pixel Module. These attributes are used by the Ultrasound, Ultrasound (Retired), and Secondary Capture Image Storage SOP Classes. Attributes not listed are not used.

Table 3-12: Image Pixel Module Attributes

Attribute Name	Tag	Type	Attribute Description
Samples per Pixel	(0028,0002)	1	MONOCHROME2=1, RGB=3, YBR_FULL_422 = 3
Photometric Interpretation	(0028,0004)	1	Configurable in DICOM Setup mode. Valid settings defined by archiver type and Transfer Syntax being used. MONOCHROME2, RGB or YBR_FULL_422
Rows	(0028,0010)	1	480
Columns	(0028,0011)	1	640
Bits Allocated	(0028,0100)	1	8
Bits Stored	(0028,0101)	1	8
High Bit	(0028,0102)	1	7
Pixel Representation	(0028,0103)	1	0
Pixel Data	(7FE0,0010)	1	Used – Basic Offset Table is set to zero length for encapsulated multi-frame images.
Planar Configuration	(0028,0006)	1C	0=Color-by-pixel - Only sent for RGB and YBR_FULL_422 images

US Region Calibration Module

Table 3-13 specifies the attributes used from the US Region Calibration Module. These attributes are used by Ultrasound, and Ultrasound (Retired) Image Storage SOP instances created by the NanoMAXX system. Attributes not listed are not used.

Table 3-13: US Region Calibration Attributes

Attribute Name	Tag	Type	Attribute Description
Sequence of Ultrasound Regions	(0018,6011)	1	Used
>Region Location Min x_0	(0018,6018)	1	Automatically generated
>Region Location Min y_0	(0018,601A)	1	Automatically generated
>Region Location Max x_1	(0018,601C)	1	Automatically generated
>Region Location Max y_1	(0018,601E)	1	Automatically generated
>Physical Units X Direction	(0018,6024)	1	
>Physical Units Y Direction	(0018,6026)	1	Automatically generated
>Physical Delta X	(0018,602C)	1	Automatically generated
>Physical Delta Y	(0018,602E)	1	Automatically generated
>Reference Pixel x_0	(0018,6020)	3	Only sent in Spectral Doppler regions.
>Reference Pixel y_0	(0018,6022)	3	Only sent in Spectral Doppler regions.

Table 3-13: US Region Calibration Attributes

Attribute Name	Tag	Type	Attribute Description
>Ref. Pixel Physical Value X	(0018,6028)	3	Only sent in Spectral Doppler regions.
>Ref. Pixel Physical Value Y	(0018,602A)	3	Only sent in Spectral Doppler regions.
>Region Spatial Format	(0018,6012)	1	Automatically generated
>Region Data Type	(0018,6014)	1	Automatically generated
>Region Flags	(0018,6016)	1	Automatically generated

US Image Module

Table 3-14 specifies the attributes used from the US Image Module. These attributes are used by Ultrasound, and Ultrasound (Retired) Image Storage SOP instances created by the NanoMAXX system. Attributes not listed are not sent.

Table 3-14: US Image Module Attributes

Attribute Name	Tag	Type	Attribute Description
Samples Per Pixel	(0028,0002)	1	MONOCHROME2=1, RGB=3, YBR_FULL_422 = 3
Photometric Interpretation	(0028,0004)	1	Configurable in DICOM Setup mode Valid settings defined by archiver type and Transfer Syntax being used. MONOCHROME2, RGB or YBR_FULL_422
Bits Allocated	(0028,0100)	1	8
Bits Stored	(0028,0101)	1	8
High Bit	(0028,0102)	1	7
Planar Configuration	(0028,0006)	1C	0=Color-by-pixel - Only sent for RGB and YBR_FULL_422 images
Pixel Representation	(0028,0103)	1	0
Image Type	(0008,0008)	2	RGB = "ORIGINAL\PRIMARY\<ExamType>\nnnn" YBR_FULL_422 or MONOCHROME2 = "DERIVED\PRIMARY\<ExamType>\nnnn" Possible values for <ExamType>: <ul style="list-style-type: none"> ▶ ABDOMINAL ▶ CHEST ▶ ENDORECTAL ▶ EPICARDIAL ▶ GYNECOLOGY ▶ BREAST ▶ ENDOCAVITARY ▶ ENDOVAGINAL ▶ FETAL HEART ▶ HEPATIC

Table 3-14: US Image Module Attributes

Attribute Name	Tag	Type	Attribute Description
			<ul style="list-style-type: none"> ▶ IMT ▶ INTRACARDIAC ▶ INTRAOPERATIVE ▶ INTRAVASCULAR ▶ MUSCULOSKELETAL ▶ NEONATAL HEAD ▶ NERVE ▶ OBSTETRICAL ▶ OPHTHALMIC ▶ ORBITAL ▶ PEDIATRIC ▶ PELVIC ▶ RETROPERITONEAL ▶ SCROTAL ▶ SMALL PARTS ▶ SUPERFICIAL ▶ TEE ▶ THYROID ▶ TRANSCRANIAL ▶ TTE ▶ US BIOPSY ▶ VASCULAR ▶ VENOUS <p>nnnn=bit map designating the image mode:</p> <ul style="list-style-type: none"> ▶ 0001 = 2D Imaging ▶ 0002 = M-Mode ▶ 0004 = CW Doppler ▶ 0008 = PW Doppler ▶ 0010 = Color Doppler ▶ 0100 = Color Power Mode
Lossy Image Compression	(0028,2110)	1C	01=Lossy Compressed - Only sent for MONOCHROME2 and JPEG Lossy Compressed images.
Ultrasound Color Data Present	(0028,0014)	3	00=Color data not present in image 01=Color data is present in image Not sent with Ultrasound (Retired) Images.
Heart Rate	(0018,1088)	3	
Transducer Data	(0018,5010)	3	

VOI LUT Module

Table 3-15 specifies the attributes used from the VOI LUT Module. These attributes are used by the Ultrasound, Ultrasound (Retired), and Secondary Capture Image Storage SOP Classes. Attributes not listed are not used.

Table 3-15: VOI LUT Module Attributes

Attribute Name	Tag	Type	Attribute Description
Window Center	(0028,1050)	3	128 - Only sent with Monochrome2
Window Width	(0028,1051)	1C	256 - Only sent with Monochrome2

SOP Common Module

Table 3-16 specifies the attributes used from the SOP Common Module. These attributes are used by the Ultrasound, Ultrasound (Retired), and Secondary Capture Image Storage SOP Classes. Attributes not listed are not used.

Table 3-16: SOP Common Module Attributes

Attribute Name	Tag	Type	Attribute Description
SOP Class UID	(0008,0016)	1	Automatically generated
SOP Instance UID	(0008,0018)	1	Automatically generated
Specific Character Set	(0008,0005)	1C	ISO_IR 100
Instance Number	(0020,0013)	3	A number that identifies this image

Store AE Behavior to C Store Status

Table 3-17 describes the behavior for C-Store response status returned from the Storage SCP. All image SOP classes supported by the Store AE exhibit the same behavior.

Table 3-17: Store AE Behavior to C-Store Status

Service Status	Further Meaning	Status Codes	Store AE Behavior
Success		0000	Continue without user notification
Refused	Out of Resources	A7xx	Association terminated. User notified.
Error	Data Set does not match SOP Class	A9xx	Association terminated. User notified.
	Cannot understand	Cxxx	Association terminated. User notified.
Warning	Coercion of data elements	B000	Ignored - Message logged.
	Data set does not match SOP class	B007	Ignored - Message logged.
	Elements discarded	B006	Ignored - Message logged.

Modality Worklist AE Specification

The Modality Worklist AE provides conformance to the following DICOM V3.0 SOP Classes as an SCU:

Table 3-18: Modality Worklist AE SOP Class Support

SOP Class Name	SOP Class UID	Conformance Level
Verification	1.2.840.10008.1.1	Standard
Modality Worklist Information Model -FIND	1.2.840.10008.5.1.4.31	Standard

Association Establishment Practices

The Modality Worklist AE will initiate an association to a device in response to the following real-world activities:

- ▶ The user initiates a manual Update Worklist (Broad Query).
- ▶ The user initiates a specific Worklist Query (Patient Based Query).
- ▶ The system initiates an Automatic Worklist Query (Broad Query).

In all cases a C-FIND command is issued to the Modality Worklist server. After the requested data is returned, the association is closed.

General

Maximum PDU size offered to SCP: 32,768 bytes

This is the maximum PDU size the Modality Worklist AE can receive and is the value offered for the maximum PDU size in the Association Request command. Once the Association is open, if the Modality Worklist AE receives a PDU that is larger than this value then the Association will be aborted.

Minimum PDU size accepted from SCP: 1,024 bytes

This is the minimum PDU size the Modality Worklist AE can be configured to send. If the Modality Worklist AE receives a maximum PDU size in the Association Accept response that is smaller than this value then the Association will be aborted immediately.

Maximum PDU size sent by SCU: 32,768 bytes

This is the maximum PDU size the Modality Worklist AE can be configured to send. The maximum PDU size sent on any Modality Worklist AE Association will be the smaller of the configured value and the maximum PDU size received in the Association Accept response.

Number of Associations

Number of simultaneous associations for the Modality Worklist AE: 1

Asynchronous Nature

The Modality Worklist AE will not use asynchronous operations.

Implementation Identifying Information

Implementation Class UID: "1.2.840.114340.3"

Implementation Version name: "Tiller_SV500"

Note

"114340" is registered by SonoSite with ANSI. Version name will be used initially as shown, but is subject to change with new versions of the DICOM capable application software.

Association Initiation by Real-World Activity

The Modality Worklist AE will open associations to the configured Worklist SCP in response to the following real-world activities: Update Worklist, Query Worklist, Automatic Worklist Query, and Get Status.

Association Initiation by: Update Worklist

The Update Worklist real-world activity initiated in Worklist screen will cause the Modality Worklist AE to open an association with the Worklist SCP, configured in the current Location.

Association Initiation by: Worklist Query

The Query Worklist command real-world activity initiated in Patient Setup screen will cause the Modality Worklist AE to open an association with the Worklist SCP, configured in the current Location.

Association Initiation by: Automatic Worklist Query

The Automatic Worklist Query real-world activity initiated by the system at periodic intervals will cause the Modality Worklist AE to open an association with the Worklist SCP, configured in the current Location.

Association Initiation by: Get Status

The Get Status real-world activity will cause the Modality Worklist AE to open an association to the Modality Worklist SCP configured in the current Location.

Proposed Presentation Contexts to a Worklist Server

Table 3-19: Modality Worklist AE Proposed Presentation Contexts to a Worklist Server

Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Extended Negotiation
Name	UID	Name List	UID List		
Verification SOP Class	1.2.840.10008.1.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None
Modality Worklist Information Model - FIND	1.2.840.10008.5.1.4.3.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None

Modality Worklist Information Model - FIND SOP Class

The Modality Worklist AE provides standard conformance to the Modality Worklist Information Model - FIND SOP Class as an SCU.

Verification SOP Class

The Modality Worklist AE provides standard conformance to the Verification SOP Class as an SCU.

Modality Worklist Attributes

Broad Worklist Query Matching Key Attributes

Table 3-20 specifies the Matching Key attributes used by Automatic Worklist C-FIND requests and manual Update Worklist C-FIND requests for Broad queries initiated by the user from the Worklist screen.

Table 3-20: Broad Worklist Query Matching Key Attributes

Attribute Name	Tag	Type	Notes
Modality	(0008,0060)	R	Selectable from list provided by User Interface
Scheduled Station AE-Title	(0040,0001)	R	Configurable – Sent as either : <ul style="list-style-type: none"> ▶ NanoMaxx’s AE Title ▶ Universal Matching
Scheduled Procedure Step Start Date	(0040,0002)	R	Configurable – Sent as either: <ul style="list-style-type: none"> ▶ Today’s date ▶ Yesterday, Today and Tomorrow date range ▶ Universal Matching

Patient Based Query Matching Key Attributes

Table 3-21 specifies the Matching Key attributes used for Worklist C-FIND requests for Patient Based queries initiated by the user from the Patient Setup screen.

Table 3-21: Patient Based Query Matching Key Attributes

Attribute Name	Tag	Type	Notes
Modality	(0008,0060)	R	Selectable from list provided by UserInterface
Scheduled Station AE-Title	(0040,0001)	R	Configurable – Sent as either : <ul style="list-style-type: none"> ▶ NanoMaxx’s AE Title ▶ Universal Matching
Scheduled Procedure Step Start Date	(0040,0002)	R	Configurable – Sent as either: <ul style="list-style-type: none"> ▶ Today’s date ▶ Yesterday, Today and Tomorrow date range ▶ Universal Matching
Patient's Name	(0010,0010)	R	Entered on Patient Setup screen. A wild card is appended to Last, First and Middle name component.
Patient ID	(0010,0020)	R	Entered on Patient Setup screen. Single value matching only.
Accession Number	(0008,0050)	O	Entered on Patient Setup screen.
Requested Procedure ID	(0040,1001)	O	Entered on Patient Setup screen.

Return Key Attributes

Table 3-22 specifies the Return Key attributes that are included in all Worklist C-FIND requests.

Table 3-22: Return Key Attributes

Attribute Name	Tag	Type	Notes
Study Instance UID	(0020,000D)	1	
Accession Number*	(0008,0050)	2	Displayed on Patient Setup screen
Referring Physician's Name	(0008,0090)	2	Displayed on Patient Setup screen
Patient's Name*	(0010,0010)	1	Displayed on Patient Setup screen. All 5 name components are preserved but only Last, First and Middle name components are displayed.
Patient ID*	(0010,0020)	1	Displayed on Patient Setup screen
Patients Birth Date	(0010,0030)	2	Displayed on Patient Setup screen
Patient's Sex	(0010,0040)	2	Displayed on Patient Setup screen
Other Patient IDs	(0010,1000)	3	

Table 3-22: Return Key Attributes

Attribute Name	Tag	Type	Notes
Additional Patient History	(0010,21B0)	3	Displayed on Patient Setup screen as Indications
Admitting Diagnoses Description	(0008,1080)	3	Displayed on Patient Setup screen as Indications if Additional Patient History is not returned.
Last Menstrual Date	(0010,21D0)	2	Displayed on Patient Setup screen with OB/GYN exam type only.
Scheduled Procedure Step Sequence	(0040,0100)	1	
>Modality	(0008,0060)	1	
>Scheduled Station AE Title	(0040,0001)	1	
>Scheduled Procedure Step Start Date	(0040,0002)	1	
>Scheduled Procedure Step Start Time	(0040,0003)	1	
>Scheduled Procedure Step Description*	(0040,0007)	1C	
>Scheduled Protocol Code Sequence	(0040,0008)	1C	
>>Code Value	(0008,0100)	1C	
>>Coding Scheme Designator	(0008,0102)	1C	
>>Code Meaning	(0008,0104)	3	
>Scheduled Procedure Step ID	(0040,0009)	1	
Requested Procedure ID*	(0040,1001)	1	Displayed on Patient Setup screen.
Requested Procedure Description*	(0032,1060)	1C	
Requested Procedure Code Sequence	(0032,1064)	1C	
>Code Value	(0008,0100)	1C	
>Coding Scheme Designator	(0008,0102)	1C	
>Code Meaning*	(0008,0104)	3	
Referenced Study Sequence	(0008,1110)	2	
>Referenced SOP Class UID	(0008,1150)	1C	
>Referenced SOP Instance UID	(0008,1155)	1C	

*This attribute from Worklist may be truncated when displayed in the NanoMAXX User Interface. However, the value contained in the attribute is preserved in full fidelity.

Worklist AE Behavior to C-FIND Status

Table 3-23 specifies the response status codes, which an SCP may return following the SCU's C-FIND request, along with the Worklist AE's associated behavior. Only those status responses that indicate some form of error condition are presented to the user. Related fields are not used.

Table 3-23: Worklist AE Behavior to C-FIND Status

Service Status	Further Meaning	Status Codes	Worklist AE Behavior
Refused	Out of resources	A700	The association is terminated. The user is notified of the failure.
Failed	Identifier does not match SOP Class	A900	The association is terminated. The user is notified of the failure.
	Unable to process	Cxxx	The association is terminated. The user is notified of the failure.
Cancel	Matching terminated due to Cancel request	FE00	The association is terminated. The user is notified that the query was incomplete.
Success	Matching is complete - No final Identifier is supplied.	0000	The Modality Worklist AE will continue operation without user notification.
Pending	Matches are continuing - Current Match is supplied and any Optional Keys were supported in the same manner as Required Keys.	FF00	The Modality Worklist AE will continue operation without user notification.
	Matches are continuing - Warning that one or more Optional Keys were not supported for existence for this Identifier.	FF01	The Modality Worklist AE will continue operation without user notification.

Worklist AE Behavior to C-FIND Status

Table 3-24 specifies the response status codes, which an SCP may return following the SCU's C-FIND request, along with the Worklist AE's associated behavior. Only those status responses that indicate some form of error condition are presented to the user. Related fields are not used.

Table 3-24: Worklist AE Behavior to C-FIND Status

Service Status	Further Meaning	Status Codes	Worklist AE Behavior
Refused	Out of resources	A700	The association is terminated. The user is notified of the failure.
Failed	Identifier does not match SOP Class	A900	The association is terminated. The user is notified of the failure.

Table 3-24: Worklist AE Behavior to C-FIND Status

Service Status	Further Meaning	Status Codes	Worklist AE Behavior
Failed	Unable to process	Cxxx	The association is terminated. The user is notified of the failure.
Cancel	Matching terminated due to Cancel request	FE00	The association is terminated. The user is notified that the query was incomplete.
Success	Matching is complete - No final Identifier is supplied.	0000	The Modality Worklist AE will continue operation without user notification.
Pending	Matches are continuing - Current Match is supplied and any Optional Keys were supported in the same manner as Required Keys.	FF00	The Modality Worklist AE will continue operation without user notification.
	Matches are continuing - Warning that one or more Optional Keys were not supported for existence for this Identifier.	FF01	The Modality Worklist AE will continue operation without user notification.

Media Export AE-Specification

This section of the conformance statement specifies the NanoMAXX compliance to DICOM Media Storage. It details the roles supported by this product.

NanoMAXX is able to export DICOM images to removable USB media stick memory. Any reference to USB in this document refers to “USB media stick memory”.

Implementation Model

The Media Export AE saves single and multi-frame US images to a USB storage device. It is associated with the local real-world activity “Export to USB”. “Export to USB” is performed upon user request for selected patient series.

Application Data Flow

Figure 3-1 Media Export Data Flow



Functional Definition of the AE

NanoMAXX can perform the following functions:

- ▶ Create a new DICOM file-set on the USB medium
- ▶ Add to an existing DICOM file-set previously created by the NanoMAXX system

Sequencing of Real-World Activities

Not applicable.

File Meta Information Options (see PS 3.10)

The implementation information written to the File Meta Header in each file is:

Implementation Class UID: "1.2.840.114340.3"

Implementation Version name: "Tiller_SV500"

Note

"114340" is registered by SonoSite with ANSI. Version name will be used initially as shown, but is subject to change with new versions of the DICOM capable application software.

AE Specifications

File Meta Information for the Application Entity

The Source Application Entity Title included in the file header is configurable. The default value set in the File Meta Information for this AE is: "DICOM Media".

Real-World Activities

Real-World Activity – "Export to USB"

"Export to USB" saves the selected DICOM SOP instances to the USB medium and creates a DICOM File Set. If a DICOM File Set created by the NanoMAXX exists on the medium, any new files selected for export will be added to the existing files. The Media Export AE acts as a File Set Creator when requested to export SOP instances from the internal storage to a USB medium. If there is insufficient space on the medium, the user will be prompted with an informative message.

Limitations: The user cannot review or manipulate DICOM files written to the USB medium on the system.

Media Storage Application Profile for the real-world activity "Export to USB"

Not applicable [FUTURE]

Options

This Application Entity supports the SOP Classes and Transfer Syntaxes listed below in [Table 3-25](#):

Table 3-25: SOP Classes and Transfer Syntaxes for Media Export

Abstract Syntax		Transfer Syntax	
Name	UID	Name List	UID List
Media Storage Directory Storage	1.2.840.10008.1.3.10	Explicit VR Little Endian	1.2.840.10008.1.2.1
Ultrasound Image Storage	1.2.840.10008.5.1.4.1. 1.6.1	Explicit VR Little Endian JPEG Baseline (Process 1)	1.2.840.10008.1.2.1 1.2.840.10008.1.2.4.50
Secondary Capture ImageStorage	1.2.840.10008.5.1.4.1. 1.7	Explicit VR Little Endian JPEG Baseline (Process 1)	1.2.840.10008.1.2.1 1.2.840.10008.1.2.4.50

[Common Composite Image IOD Modules](#) on page 15 describes image module usage by NanoMAXX.

Augmented and Private Application Profiles

Not applicable.

Media Configuration


The Application Entity Titles configurable for Media Services are listed below:

- ▶ Application Entity: "Media Export"
- ▶ Default AE Title: "DICOM Media"

Media Storage SOP Class

The following diagram illustrates the relationship between directory entities in the Basic Directory module produced by NanoMAXX. It is based on the basic DICOM entity relationship model.

Figure 3-2 NanoMAXX Directory Entity Relationship Diagram



The Media Storage SOP Class uses the Basic Directory IOD Modules as shown in [Table 3-26](#).

Table 3-26: Basic Directory IOD Modules

Module	Reference	Usage
File-set Identification	3.1.4.1	M
Directory Information	3.4.7.2	U

Information Module Definitions

File-Set Identification Module

[Table 3-27](#) specifies the attributes used from the File-set Identification Module.

Table 3-27: File-Set Identification Module

Attribute Name	Tag	Type	Attribute Description
File-set ID	(0004,1130)	2	“SONO_EXPORT”
File-set Descriptor ID	(0004,1141)	3	Not Used
Specific Character Set of File- set Descriptor File	(0004,1142)	1C	Not Used

Directory Information Module

Table 3-28 specifies the attributes used from the Directory Information Module.

Table 3-28: Directory Information Module

Attribute Name	Tag	Type	Attribute Description
Offset of the First Directory Record of the Root Directory Entity	(0004,1200)	1	See PS 3.3
Offset of the Last Directory Record of the Root Directory Entity	(0004,1202)	1	See PS 3.3
File-set Consistency Flag	(0004,1212)	1	See PS 3.3
Directory Record Sequence	(0004,1220)	2	See PS 3.3
>Offset of the Next Directory Record	(0004,1400)	1C	See PS 3.3
>Record In-use Flag	(0004,1410)	1C	NanoMAXX sets all created records to 0xFFFF
>Offset of Referenced Lower-Level Directory Entity	(0004,1420)	1C	See PS 3.3
>Directory Record Type	(0004,1430)	1C	NanoMAXX Supported Values: PATIENT, STUDY, SERIES, IMAGE
>Referenced File ID	(0004,1500)	1C	See PS 3.3
>Referenced SOP Class UID in File	(0004,1510)	1C	See PS 3.3
>Referenced SOP Instance UID in File	(0004,1511)	1C	See PS 3.3
>Referenced Transfer Syntax in UID in File	(0004,1512)	1C	See PS 3.3

Patient Keys

Table 3-29 specifies the additional keys used for directory records of type PATIENT.

Table 3-29: Patient Keys

Attribute Name	Tag	Type	Attribute Description
Patient's Name	(0010,0010)	2	Reference 3.2.4.1
Patient ID	(0010,0020)	1	Reference 3.2.4.1

Study Keys

Table 3-30 specifies the additional keys used for directory records of type STUDY.

Table 3-30: Study Keys

Attribute Name	Tag	Type	Attribute Description
Study Date	(0008,0020)	1	Reference 3.2.4.2
Study Time	(0008,0030)	1	Reference 3.2.4.2
Study Description	(0008,1030)	2	Reference 3.2.4.2
Study Instance UID	(0020,000D)	1C	Reference 3.2.4.2
Study ID	(0020,0010)	1	Reference 3.2.4.2
Accession Number	(0008,0050)	2	Reference 3.2.4.2

Series Keys

Table 3-31 specifies the additional keys used for directory records of type SERIES.

Table 3-31: Series Keys

Attribute Name	Tag	Type	Attribute Description
Modality	(0008,0060)	1	Reference 3.2.4.4
Series Instance UID	(0020,000E)	1	Reference 3.2.4.4
Series Number	(0020,0011)	1	Reference 3.2.4.4

Image Keys

Table 3-32 specifies the additional keys used for directory records of type IMAGE.

Table 3-32: Image Keys

Attribute Name	Tag	Type	Attribute Description
Instance Number	(0020,0013)	1	Reference 3.2.4.7

CHAPTER 4

Communication Profiles

TCP/IP Stack

The TCP/IP protocol is used.

Extensions/Specializations/Privatizations

Standard Extended/Specialized/Private SOPs

SonoSite uses the tag range of (0019,XXXX) for private tags in US Image Storage objects. These tags include additional report and image metadata intended for use by SonoSite applications, and are included only when the “Include private tags” option is selected on the system.

Table 4-1: Private Tags

Tag	VR	Attribute Description
(0019,0010)	LO	Private Data
(0019,1010)	UT	Private Data (only present in first image in series)
(0019,1020)	UT	Private Data (only present in first image in series)
(0019,1030)	UT	Private Data (only present in first image in series)
(0019,1040)	UT	Private Data (only present in first image in series)
(0019,1050)	UT	Private Data
(0019,1060)	UT	Private Data

Private Transfer Syntaxes

None

CHAPTER 6

Configuration

AE Title/Presentation Address Mapping

The NanoMAXX AE Title and networking parameters are configurable in DICOM Setup Mode. Port number 104 is the default used for DICOM communication.

Configurable Parameters

NanoMAXX Configurable Parameters per Network Location

The NanoMAXX system can be configured to operate in multiple network locations. The NanoMAXX local device settings and remote device settings (e.g. Archivers/Worklist) can be configured for each location. These parameters are intended to be configured by a network/DICOM administrator.

Configurable NanoMAXX Networking and DICOM parameters:

- ▶ DHCP (default = disabled)
- ▶ Hostname (Name field)
- ▶ DICOM AE Title
- ▶ IP Address (disabled if DHCP is selected)
- ▶ Subnet Mask (disabled if DHCP is selected)
- ▶ Default Gateway (disabled if DHCP is selected)
- ▶ Network Write Timeout
- ▶ Network Read Timeout
- ▶ Network speed (Auto, 100Mb/10Mb, Full/Half duplex)
- ▶ Transfer Images (End of exam, During the exam)
- ▶ Port (default = 104)
- ▶ Wireless properties (see NanoMAXX User Guide for detailed configuration information)

Configurable Parameters per Remote Device Instance

Every archiver and Modality Worklist device that NanoMAXX is setup to communicate with has a set of parameters that are configurable in Setup mode. These parameters are intended to be configured by a network/DICOM administrator.

Configurable parameters for each device instance:

- ▶ DICOM AE Title
- ▶ Hostname (Name field)
- ▶ IP Address
- ▶ Port Number

Configurable parameters for each Archiver device instance:

- ▶ SOP Class Ultrasound / Ultrasound Retired / Secondary Capture
- ▶ Photometric Interpretation Monochrome2 / RGB/ YBR_FULL_422 (used for JPEG Baseline)
- ▶ Transfer Syntax ELE/ILE or JPEG Baseline
- ▶ Send Images Only Selection
- ▶ Max Retries
- ▶ Retry Interval
- ▶ Inclusion of private tags

Configurable parameters for the Worklist SCP device instance:

- ▶ Automatic Query Enable On/Off
- ▶ Automatic Query Interval Selection = 30 minutes, 1, 2, 4, 8, 12, 24 hours
- ▶ Automatic Query Start Time Selection = 0:00 to 23:00 hours
- ▶ Modality Selectable from list in user interface
- ▶ Scheduled Station AE Title This NanoMAXX system only or universal matching (used for Broad & Patient Based Queries)
- ▶ SPS Date Today; "Yesterday, Today & Tomorrow"; or universal matching (used for Broad & Patient Based Queries)

Other Configurable Parameters

These settings apply independent of network configuration:

- ▶ Photometric Interpretation Monochrome2 / RGB / YBR_FULL_422 (removable media)

CHAPTER 7

Support of Extended Character Sets

The NanoMAXX system supports the ISO 8859 Latin 1 (ISO-IR 100) character set family.

The Specific Character Set key attribute (0008,0005), a type 1C attribute, may be returned by an SCP if that device supports any character set encodings beyond the ISO_IR 6. If the tag is not present in the Worklist query result, the default (i.e. ISO_IR 6, i.e. ASCII) is assumed. If the tag is present, only ISO_IR 6 (ASCII) or ISO_IR 100 (Latin Alphabet # 1) are supported by the NanoMAXX system. All other character set encodings are unsupported and will cause the system to issue a C-Find Cancel. All query results data acquired up to the first detection of an unsupported character set encoding are retained and presented to the user.

FUJIFILM
Value from Innovation

SonoSite

D07914